

Flamingo (bath frame)		
User information appropriate for use by people with visual, reading or cognitive disability	The latest version of all instructions is always available on the R82 website support.R82.org and can be printed in larger sizes	
Information available in pictogram	The instructions are made as pictograms	
Intended use and intended environment including standard options	The Flamingo Bath Frame is a toilet/bathing frame with user friendly angle and height adjustment. The Flamingo Bath Frame is designed for Flamingo seats and can be placed over the toilet or used freestanding with a pan. The frame comes in four sizes and is only for indoor use but can in an emergency be taken outside.	
Maintenance instructions	Maintenance of a medical device is wholly the responsibility of the owner of that device. Failure to maintain a device in accordance with the instructions may invalidate the device's warranty. Furthermore, failure to maintain a device may compromise the clinical condition or safety of users and/or their caregivers.	
	_	5
	Frequency 1 st day	Read the User Guide carefully before use and store it in a safe place for future reference Store the enclosed tools on/at the product
	After every use	- Wipe the product dry
	Daily	- Use a dry cloth to clean the product
	·	 Visual inspection. Repair or replace any damaged or worn-out parts
		 Ensure that all fasteners are in place and fitted correctly
		 Check all fittings and buckles for signs of worn out parts
		 Be aware of any MAX mark indications before adjustment
	Weekly	 Check that all castors are moving freely and that all wheel locks/brakes are easily activated
		 Wipe off any residue and dirt from the product using a cloth with warm water and a mild detergent/soap and dry before use
	Monthly	- Nuts and bolts on the product are tightened
	Appually	- Oil the swing-away parts
	Annually	 Inspect the frame for cracks or signs for worn- out parts and carry out annual service. Never use a product which appears to be substandard or faulty
Cleaning description	This product can be washed at 85° using a mild detergent for 3 minutes in a washing machine designed for medical devices. Use the machine's drying function to dry the product.	
	This product can be washed at 60° using a mild detergent for 10 minutes in a washing machine designed for medical devices. Use the machine's drying function to dry the product.	
	Before washing Before washing remove the following items and clean them separately: Accessories and Seat	
Disinfection description	The product can be disinfected with a 70% disinfectant IPA solution. It is recommended to wipe off any residue and dirt from the product, using a	



Flamingo (bath frame)		
	cloth with warm water and a mild detergent/soap without chlorine and let it dry before disinfection.	
Dimensions (Incl. max load)	http://www.R82.org - Select product or click here	
Mass	N/A	
Used in combination with other products	N/A	
Warning about dangerous combinations	The CE mark must be removed when rebuilding the product, when used in combination with a product from another manufacturer or when using other than original R82 spare parts and fittings.	
List of accessories, detachable parts and materials	http://www.R82.org - Select product or click here	
Programmable controller information	N/A	
Operator control adjustments	Manual	
Folding and dismantling	N/A	
Transport instructions (unoccupied)	Land and/or air	
Sound power level	N/A	
Turning diameter	NA	
Forces on the wall	N/A	
Electromagnetic environment warning	N/A	
Electromagnetic influence warning	N/A	
Life-time	The lifetime of the product in normal use is 5 years	

Page 2 of 2